



DOEACC SOCIETY – GORAKHPUR CENTRE

(A Scientific Body of Dept. of Information Technology, Ministry of Comm. & Information Technology, Govt. of India)

M.M.M.Engg. College Campus, Gorakhpur-273010(U.P.)

Phone: 0551- 2273371, Fax: 2273873

Website: www.doeaccgkp.edu.in

CDS/GP/7.4.1/F08/R0

Ref No.: DOEACC/Gkp/224/15/09-10

Date 16-12-2009

INVITATION TO QUOTE RATES

(Last date for Submission : **05-01-2010**)

To,

On website copy

Dear Sir,

You are requested to quote rate (s) for the items mentioned below as per the specifications given below.

Sl.no.	Name of the Item(s)	Description / Specification	Qty. Reqd.	Unit
01	100 MHz Digital Storage CRO	Enclosed Annexure-I	01	No.

The Terms & Conditions as per enclosure should be read carefully and noted in order to comply with.

Yours faithfully,

For DOEACC Centre, Gorakhpur
Authorised Signatory
PTO

TERMS AND CONDITIONS (FOR LIMITED TENDER)

1. The quotation should be addressed to Director, DOEACC SOCIETY-GORAKHPUR CENTRE.
2. The price quoted should be firm and F.O.R. DOEACC SOCIETY-GORAKHPUR CENTRE. Wherever possible, educational prices should be quoted. Quantity indicated is tentative and is subject to change.
3. The quotation/offer should be submitted on or before last date of submission (Due Date) in a sealed envelope. The envelope should be superscribed as “**Quotation for “100 MHz Digital CRO ”**”and should bear the enquiry number and due date on the top. Quotations received after the due date will not be considered.
4. Quotations will not be accepted by Fax, Email or any such electronic data transfer form.
5. Customs Duty, Excise Duty, Sales Tax, C & F, Packing & Forwarding, Insurance, Transportation, Octroi, Installation, Training, etc charges should be separately mentioned. Discount, if any, should also be mentioned. Total price of the goods should also be mentioned.
6. Octroi of Respective Municipal Corporation may either be included or may be shown as ‘at actual’.
7. Delivery period and place of delivery should be clearly mentioned. Delivery Location should be F.O.R. DOEACC SOCIETY-GORAKHPUR CENTRE..
8. The goods should be quoted along with a guarantee/warranty period of minimum ‘one year’ from the date of successful installation.
9. Validity of the quotation should be mentioned. Normally quotations should be valid for at least 60 days from the date of opening of the quotation.
10. The bidding firm must have a valid Sales Tax Registration Number which must be indicated on the quotation.
11. Payment Terms:
 - a. No advance payment (Prepayment) will be made at any cost.
 - b. Payment @ 90 % of the invoice value to the successful bidder will be made after physical verification / satisfactory installation and commissioning of the goods at DOEACC SOCIETY-GORAKHPUR CENTRE., Gorakhpur will make remaining 10% payments after completion of satisfactory training / acceptance of the goods.
12. Detailed description & technical specifications of the goods quoted by you should be provided. Technical literature / brochure should accompany the quotation.
13. Manufacturing license or authorized Distributor / Principal or Dealer Certificate copy should be enclosed.
14. OEM items if any, should be supplied in the original sealed packing of the original equipment manufacturer and should be opened at SOCIETY-GORAKHPUR CENTRE.in the presence of Center’s representative only.
15. It will be binding on the part of the successful bidder to supply the goods at the rates quoted, failing to which the EMD will be forfeited the firm will be removed from the suppliers list and no further enquiries would be sent.
16. DOEACC SOCIETY-GORAKHPUR CENTRE. does not bind itself to accept the lowest or any such quotation and has the right to accept or reject whole or any part of tenders or a portion of the supply of goods without assigning any reasons. No correspondence in case of rejected tenders will be entertained. Incomplete quotations are liable to be rejected.
17. All questions, disputes or differences arising under and out of or in connection with the contract, if concluded, shall be referred to the sole arbitrator appointed by the Director, DOEACC SOCIETY-GORAKHPUR CENTRE.

-----XXXXXXXXXX-----

100 MHz Digital Storage Oscilloscope

Technical Specification:

Real time Band Width	:DC- 100 MHz on all Channels
No. of Channels	: Two with Externalchannels
Sample Rate (Max)	: 1 GS/s on all channels
Record length	: 2500 Points per channel
Vertical Resolution	: 8 Bits
Vertical sensitivity	: 5mV to 5 Volts
Time Base Range	: 5 ns to 50 sec per division
Time Base accuracy	: 50 PPM.
Sweep Mode	: Auto, Normal and Single Sequerce
Input Coupling	: AC, DC Ground
Max Input Voltage	: 300V CAT II
Triggering	: Edge, Pulse width
Analysis Mode	: Average Normal, Peak detect and Roll mode
Measurements	:Automatic measurement, time & Voltage cursors math (Add, sub) and FFT.
Display	: LCD Monochrome Display
Format	: Y-T and X-Y
Power suppldsy	: 230V 50 HzAC.
Acessories	: One 1x, 10x switchable probe per channel and manual/Technical Manual etc.

-----XXXXXXXXXXXXXXXXXX-----